



EC Declaration of Conformity

We,

Labotron Instruments Ltd.
10 HSIDC Industrial Estates
Ambala Cantt – 133 006
Haryana, INDIA

Declare under sole responsibility that the following described product in our delivered version complies with the appropriate basic safety and health requirements of the **EC Low Voltage Directive (73/23/EEC & amended as 2006/95/EC)** and **EC Electromagnetic Compatibility Directive (89/336/EEC & amended as 2004/108/EC)** based on its design and type, as brought into circulation by us. In case of alteration of the machine, not agreed upon by us, this declaration will lose its validity.

Description of the Electrical Equipment: LIGHT (COMPOUND) MICROSCOPE
Models: CxL, CxL POL, LX400, LX400eFL, LX400P, Lx500, TCM400 and MET 400
Year of CE marking authorization September 2012

Applicable Directives and standards:

EC Low Voltage Directive (2006/95/EC)

EMC Directive (2004/108/EC)

EN 61010-1:2001

EN/IEC 61326-1:2006
EN/IEC 61000-3-3:2008
EN/IEC 61000-3-2:2006
EN55011:2007

AUTHORIZED SIGNATORY:

DATE: SEPTEMBER 18, 2012

SUKHVINDER BAL
Manager (QA & Compliance)

E-mail: sbal@labomed.in

Regd. Office :
FLAT NO. 6, GAURI APARTMENTS
3-4, RAJESH PILOT LANE
NEW DELHI - 110 011 INDIA
Phone : (011) 23018882



ISO 9001 : 2008
FILE NO. A8956